

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121421\
 Data File : VX025791.D
 Acq On : 14 Dec 2021 20:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 15 07:10:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	119	0.00
2 T	Dichlorodifluoromethane	0.447	0.446	0.2	125	0.00
3 P	Chloromethane	0.582	0.505	13.2	101	0.00
4 C	Vinyl Chloride	0.619	0.543	12.3#	107	0.00
5 T	Bromomethane	0.442	0.356	19.5	96	0.00
6 T	Chloroethane	0.449	0.370	17.6	105	0.00
7 T	Trichlorofluoromethane	1.023	0.888	13.2	111	0.00
8 T	Diethyl Ether	0.404	0.322	20.3	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.532	0.472	11.3	114	0.00
10 T	Methyl Iodide	0.678	0.618	8.8	100	0.00
11 T	Tert butyl alcohol	0.144	0.113	21.5	98	0.00
12 CM	1,1-Dichloroethene	0.529	0.458	13.4#	105	0.00
13 T	Acrolein	0.099	0.084	15.2	109	0.00
14 T	Allyl chloride	0.966	0.780	19.3	96	0.00
15 T	Acrylonitrile	0.337	0.291	13.6	98	0.00
16 T	Acetone	0.344	0.251	27.0#	85	0.00
17 T	Carbon Disulfide	1.516	1.215	19.9	99	0.00
18 T	Methyl Acetate	1.096	0.930	15.1	95	0.00
19 T	Methyl tert-butyl Ether	1.870	1.617	13.5	97	0.00
20 T	Methylene Chloride	0.699	0.535	23.5	96	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.517	12.5	102	0.00
22 T	Diisopropyl ether	1.874	1.627	13.2	98	0.00
23 T	Vinyl Acetate	1.450	1.268	12.6	95	0.00
24 P	1,1-Dichloroethane	1.070	0.908	15.1	99	0.00
25 T	2-Butanone	0.484	0.413	14.7	95	0.00
26 T	2,2-Dichloropropane	0.812	0.595	26.7#	88	0.00
27 T	cis-1,2-Dichloroethene	0.700	0.608	13.1	101	0.00
28 T	Bromochloromethane	0.482	0.428	11.2	96	0.00
29 T	Tetrahydrofuran	0.308	0.266	13.6	97	0.00
30 C	Chloroform	1.173	0.992	15.4#	99	0.00
31 T	Cyclohexane	0.875	0.820	6.3	118	0.00
32 T	1,1,1-Trichloroethane	1.021	0.866	15.2	101	0.00
33 S	1,2-Dichloroethane-d4	0.719	0.639	11.1	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.349	0.331	5.2	105	0.00
36 T	1,1-Dichloropropene	0.473	0.428	9.5	105	0.00
37 T	Ethyl Acetate	0.525	0.471	10.3	96	0.00
38 T	Carbon Tetrachloride	0.550	0.465	15.5	100	0.00
39 T	Methylcyclohexane	0.529	0.513	3.0	123	0.00
40 TM	Benzene	1.411	1.281	9.2	100	0.00
41 T	Methacrylonitrile	0.291	0.260	10.7	98	0.00
42 TM	1,2-Dichloroethane	0.536	0.471	12.1	96	0.00
43 T	Isopropyl Acetate	0.846	0.726	14.2	93	0.00
44 TM	Trichloroethene	0.393	0.365	7.1	106	0.00
45 C	1,2-Dichloropropane	0.356	0.309	13.2#	96	0.00
46 T	Dibromomethane	0.277	0.240	13.4	96	0.00
47 T	Bromodichloromethane	0.546	0.458	16.1	92	0.00
48 T	Methyl methacrylate	0.399	0.357	10.5	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025791.D
 Acq On : 14 Dec 2021 20:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 15 07:10:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	100	0.00
50 S	Toluene-d8	1.271	1.210	4.8	108	0.00
51 T	4-Methyl-2-Pentanone	0.554	0.503	9.2	98	0.00
52 CM	Toluene	0.899	0.851	5.3#	102	0.00
53 T	t-1,3-Dichloropropene	0.520	0.445	14.4	88	0.00
54 T	cis-1,3-Dichloropropene	0.569	0.499	12.3	93	0.00
55 T	1,1,2-Trichloroethane	0.385	0.345	10.4	99	0.00
56 T	Ethyl methacrylate	0.554	0.510	7.9	97	0.00
57 T	1,3-Dichloropropane	0.632	0.558	11.7	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.236	2.5	104	0.00
59 T	2-Hexanone	0.420	0.383	8.8	97	0.00
60 T	Dibromochloromethane	0.437	0.370	15.3	89	0.00
61 T	1,2-Dibromoethane	0.418	0.372	11.0	98	0.00
62 S	4-Bromofluorobenzene	0.483	0.465	3.7	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	112	0.00
64 T	Tetrachloroethene	0.399	0.378	5.3	108	0.00
65 PM	Chlorobenzene	1.042	0.957	8.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.356	12.5	95	0.00
67 C	Ethyl Benzene	1.762	1.692	4.0#	103	0.00
68 T	m/p-Xylenes	0.686	0.667	2.8	102	0.00
69 T	o-Xylene	0.675	0.657	2.7	103	0.00
70 T	Styrene	1.123	1.097	2.3	100	0.00
71 P	Bromoform	0.335	0.278	17.0	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
73 T	Isopropylbenzene	3.629	3.334	8.1	107	0.00
74 T	N-amyl acetate	1.460	1.298	11.1	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.265	1.071	15.3	97	0.00
76 T	1,2,3-Trichloropropane	1.146	0.954	16.8	97	0.00
77 T	Bromobenzene	0.931	0.850	8.7	103	0.00
78 T	n-propylbenzene	4.105	3.842	6.4	107	0.00
79 T	2-Chlorotoluene	2.610	2.299	11.9	102	0.00
80 T	1,3,5-Trimethylbenzene	3.010	2.817	6.4	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.264	28.6#	79	0.00
82 T	4-Chlorotoluene	2.923	2.610	10.7	101	0.00
83 T	tert-Butylbenzene	2.950	2.745	6.9	108	0.00
84 T	1,2,4-Trimethylbenzene	3.007	2.830	5.9	103	0.00
85 T	sec-Butylbenzene	3.546	3.498	1.4	113	0.00
86 T	p-Isopropyltoluene	2.978	2.880	3.3	109	0.00
87 T	1,3-Dichlorobenzene	1.734	1.575	9.2	103	0.00
88 T	1,4-Dichlorobenzene	1.771	1.577	11.0	102	0.00
89 T	n-Butylbenzene	2.484	2.385	4.0	111	0.00
90 T	Hexachloroethane	0.585	0.453	22.6	90	0.00
91 T	1,2-Dichlorobenzene	1.656	1.526	7.9	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.226	20.4	88	0.00
93 T	1,2,4-Trichlorobenzene	0.848	0.880	-3.8	108	0.00
94 T	Hexachlorobutadiene	0.352	0.357	-1.4	124	0.00
95 T	Naphthalene	2.820	3.021	-7.1	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025791.D
Acq On : 14 Dec 2021 20:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 15 07:10:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.846	0.889	-5.1	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6