



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025768.D
 Acq On : 14 Dec 2021 09:42
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Title : SW846 8260
 Last Update : Tue Dec 14 07:02:11 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	6566	PASS
75	95	30	60	52.4	16801	PASS
95	95	100	100	100.0	32088	PASS
96	95	5	9	6.4	2065	PASS
173	174	0.00	2	1.3	283	PASS
174	95	50	100	68.3	21905	PASS
175	174	5	9	7.9	1721	PASS
176	174	95	101	97.4	21344	PASS
177	176	5	9	7.1	1511	PASS