

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
 Data File : VX033417.D
 Acq On : 14 Dec 2022 14:21
 Operator : JC/MD
 Sample : N6049-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-121322

Quant Time: Dec 14 22:49:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201421	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37946	46.565	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.120%
35) Dibromofluoromethane	5.385	113	42548	46.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	8.647	98	87292	46.059	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.120%
62) 4-Bromofluorobenzene	11.079	95	67746	46.750	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%
Target Compounds						
16) Acetone	2.398	43	3748	6.215	ug/l	90
20) Methylene Chloride	2.794	84	2532	1.883	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
Data File : VX033417.D
Acq On : 14 Dec 2022 14:21
Operator : JC/MD
Sample : N6049-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-121322

Quant Time: Dec 14 22:49:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

