

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
 Data File : VX033421.D
 Acq On : 14 Dec 2022 15:53
 Operator : JC/MD
 Sample : PB149591ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149591ZHE#02

Quant Time: Dec 14 22:50:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204741	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36786	45.579	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.160%
35) Dibromofluoromethane	5.385	113	40159	43.609	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.220%
50) Toluene-d8	8.647	98	87678	45.513	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.020%#
62) 4-Bromofluorobenzene	11.079	95	67378	45.742	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.480%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10222	17.116	ug/l	93
20) Methylene Chloride	2.794	84	5922	4.446	ug/l	93
37) Ethyl Acetate	4.721	43	16325	9.619	ug/l	99
43) Isopropyl Acetate	6.336	43	72254	26.459	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
Data File : VX033421.D
Acq On : 14 Dec 2022 15:53
Operator : JC/MD
Sample : PB149591ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149591ZHE#02

Quant Time: Dec 14 22:50:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

