

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
 Data File : VX033425.D
 Acq On : 14 Dec 2022 17:25
 Operator : JC/MD
 Sample : N6024-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-21

Quant Time: Dec 14 22:52:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	39116	48.154	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.300%
35) Dibromofluoromethane	5.391	113	42820	46.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	91758	47.328	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.660%
62) 4-Bromofluorobenzene	11.079	95	72627	48.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.980%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	10287	17.114	ug/l	96
20) Methylene Chloride	2.788	84	6076	4.533	ug/l	89
25) 2-Butanone	4.562	43	7447	8.631	ug/l	94
37) Ethyl Acetate	4.720	43	15426	9.031	ug/l	99
43) Isopropyl Acetate	6.342	43	68089	24.776	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
Data File : VX033425.D
Acq On : 14 Dec 2022 17:25
Operator : JC/MD
Sample : N6024-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-21

Quant Time: Dec 14 22:52:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

