

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX121422\  
 Data File : VX033426.D  
 Acq On : 14 Dec 2022 17:49  
 Operator : JC/MD  
 Sample : N6048-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MH-22

Quant Time: Dec 14 22:52:43 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X120722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 07 12:43:54 2022  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |                | Response | Conc       | Units | Dev(Min) |
|-----------------------------|-----------|----------------|----------|------------|-------|----------|
| -----                       |           |                |          |            |       |          |
| Internal Standards          |           |                |          |            |       |          |
| 1) Pentafluorobenzene       | 5.556     | 168            | 121087   | 50.000     | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763     | 114            | 201992   | 50.000     | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117            | 193250   | 50.000     | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152            | 85096    | 50.000     | ug/l  | 0.00     |
| System Monitoring Compounds |           |                |          |            |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958     | 65             | 40076    | 51.590     | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range 78 - 117 | Recovery | = 103.180% |       |          |
| 35) Dibromofluoromethane    | 5.391     | 113            | 44401    | 48.872     | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range 75 - 124 | Recovery | = 97.740%  |       |          |
| 50) Toluene-d8              | 8.647     | 98             | 93040    | 48.954     | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range 92 - 112 | Recovery | = 97.900%  |       |          |
| 62) 4-Bromofluorobenzene    | 11.079    | 95             | 73769    | 50.763     | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range 83 - 123 | Recovery | = 101.520% |       |          |
| Target Compounds            |           |                |          |            |       |          |
|                             |           |                |          |            |       | Qvalue   |
| 16) Acetone                 | 2.386     | 43             | 10523    | 18.307     | ug/l  | 99       |
| 20) Methylene Chloride      | 2.794     | 84             | 7475     | 5.831      | ug/l  | 89       |
| 25) 2-Butanone              | 4.562     | 43             | 7872     | 9.541      | ug/l  | 97       |
| 37) Ethyl Acetate           | 4.721     | 43             | 15199    | 9.077      | ug/l  | 99       |
| 43) Isopropyl Acetate       | 6.342     | 43             | 67607    | 25.095     | ug/l  | 99       |
| -----                       |           |                |          |            |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX121422\  
Data File : VX033426.D  
Acq On : 14 Dec 2022 17:49  
Operator : JC/MD  
Sample : N6048-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-22

Quant Time: Dec 14 22:52:43 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X120722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 07 12:43:54 2022  
Response via : Initial Calibration

