

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
 Data File : VX033428.D
 Acq On : 14 Dec 2022 18:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 22:53:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.349	0.332	4.9	89	0.00
3 P	Chloromethane	0.404	0.344	14.9	83	0.00
4 C	Vinyl Chloride	0.461	0.408	11.5#	84	0.00
5 T	Bromomethane	0.338	0.301	10.9	88	0.00
6 T	Chloroethane	0.350	0.315	10.0	82	0.00
7 T	Trichlorofluoromethane	0.866	0.608	29.8#	66	0.00
8 T	Diethyl Ether	0.313	0.209	33.2#	67	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.459	0.331	27.9#	65	0.00
10 T	Methyl Iodide	0.503	0.374	25.6#	67	0.00
11 T	Tert butyl alcohol	0.098	0.062	36.7#	68	0.01
12 CM	1,1-Dichloroethene	0.453	0.313	30.9#	65	0.00
13 T	Acrolein	0.106	0.071	33.0#	62	0.00
14 T	Allyl chloride	0.724	0.455	37.2#	58	0.00
15 T	Acrylonitrile	0.243	0.158	35.0#	61	0.00
16 T	Acetone	0.237	0.127	46.4#	55	0.00
17 T	Carbon Disulfide	1.163	0.721	38.0#	58	0.00
18 T	Methyl Acetate	0.687	0.456	33.6#	62	0.00
19 T	Methyl tert-butyl Ether	1.570	1.059	32.5#	63	0.00
20 T	Methylene Chloride	0.529	0.343	35.2#	64	0.00
21 T	trans-1,2-Dichloroethene	0.496	0.345	30.4#	65	0.00
22 T	Diisopropyl ether	1.551	1.038	33.1#	60	0.00
23 T	Vinyl Acetate	1.204	0.817	32.1#	61	0.00
24 P	1,1-Dichloroethane	0.911	0.601	34.0#	63	0.00
25 T	2-Butanone	0.341	0.228	33.1#	62	0.00
26 T	2,2-Dichloropropane	0.670	0.463	30.9#	63	0.00
27 T	cis-1,2-Dichloroethene	0.593	0.418	29.5#	66	0.00
28 T	Bromochloromethane	0.381	0.380	0.3	89	0.00
29 T	Tetrahydrofuran	0.221	0.206	6.8	86	0.00
30 C	Chloroform	1.012	1.023	-1.1#	94	0.00
31 T	Cyclohexane	0.762	0.749	1.7	89	0.00
32 T	1,1,1-Trichloroethane	0.870	0.901	-3.6	93	0.00
33 S	1,2-Dichloroethane-d4	0.321	0.293	8.7	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.225	0.225	0.0	95	0.00
36 T	1,1-Dichloropropene	0.438	0.453	-3.4	92	0.00
37 T	Ethyl Acetate	0.414	0.408	1.4	88	0.00
38 T	Carbon Tetrachloride	0.491	0.502	-2.2	90	0.00
39 T	Methylcyclohexane	0.491	0.508	-3.5	89	0.00
40 TM	Benzene	1.278	1.323	-3.5	93	0.00
41 T	Methacrylonitrile	0.228	0.229	-0.4	87	0.00
42 TM	1,2-Dichloroethane	0.508	0.500	1.6	90	0.00
43 T	Isopropyl Acetate	0.667	0.673	-0.9	90	0.00
44 TM	Trichloroethene	0.352	0.372	-5.7	94	0.00
45 C	1,2-Dichloropropane	0.327	0.338	-3.4#	94	0.00
46 T	Dibromomethane	0.249	0.253	-1.6	93	0.00
47 T	Bromodichloromethane	0.485	0.497	-2.5	91	0.00
48 T	Methyl methacrylate	0.344	0.345	-0.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
 Data File : VX033428.D
 Acq On : 14 Dec 2022 18:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 22:53:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	90	0.00
50 S	Toluene-d8	0.470	0.470	0.0	91	0.00
51 T	4-Methyl-2-Pentanone	0.422	0.425	-0.7	88	0.00
52 CM	Toluene	0.827	0.881	-6.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.473	0.508	-7.4	93	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.557	-7.7	94	0.00
55 T	1,1,2-Trichloroethane	0.347	0.370	-6.6	95	0.00
56 T	Ethyl methacrylate	0.489	0.525	-7.4	92	0.00
57 T	1,3-Dichloropropane	0.558	0.584	-4.7	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.224	0.237	-5.8	91	0.00
59 T	2-Hexanone	0.311	0.311	0.0	87	0.00
60 T	Dibromochloromethane	0.383	0.404	-5.5	92	0.00
61 T	1,2-Dibromoethane	0.370	0.395	-6.8	94	0.00
62 S	4-Bromofluorobenzene	0.360	0.376	-4.4	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.342	0.346	-1.2	92	0.00
65 PM	Chlorobenzene	1.013	1.049	-3.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.404	-5.5	93	0.00
67 C	Ethyl Benzene	1.764	1.862	-5.6#	92	0.00
68 T	m/p-Xylenes	0.686	0.734	-7.0	93	0.00
69 T	o-Xylene	0.678	0.717	-5.8	92	0.00
70 T	Styrene	1.111	1.198	-7.8	92	0.00
71 P	Bromoform	0.303	0.315	-4.0	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.624	3.707	-2.3	92	0.00
74 T	N-amyl acetate	1.236	1.298	-5.0	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.137	1.1	92	0.00
76 T	1,2,3-Trichloropropane	1.017	1.024	-0.7	94	0.00
77 T	Bromobenzene	0.915	0.928	-1.4	94	0.00
78 T	n-propylbenzene	4.047	4.216	-4.2	92	0.00
79 T	2-Chlorotoluene	2.501	2.527	-1.0	92	0.00
80 T	1,3,5-Trimethylbenzene	3.017	3.128	-3.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.327	0.314	4.0	88	0.00
82 T	4-Chlorotoluene	2.908	2.909	-0.0	92	0.00
83 T	tert-Butylbenzene	3.008	3.135	-4.2	93	0.00
84 T	1,2,4-Trimethylbenzene	3.011	3.102	-3.0	92	0.00
85 T	sec-Butylbenzene	3.510	3.698	-5.4	92	0.00
86 T	p-Isopropyltoluene	3.026	3.151	-4.1	92	0.00
87 T	1,3-Dichlorobenzene	1.658	1.698	-2.4	93	0.00
88 T	1,4-Dichlorobenzene	1.693	1.716	-1.4	95	0.00
89 T	n-Butylbenzene	2.478	2.630	-6.1	92	0.00
90 T	Hexachloroethane	0.514	0.511	0.6	89	0.00
91 T	1,2-Dichlorobenzene	1.657	1.691	-2.1	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.230	7.3	87	0.00
93 T	1,2,4-Trichlorobenzene	1.001	1.055	-5.4	97	0.00
94 T	Hexachlorobutadiene	0.401	0.430	-7.2	99	0.00
95 T	Naphthalene	3.339	3.521	-5.5	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121422\
Data File : VX033428.D
Acq On : 14 Dec 2022 18:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 14 22:53:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	1.049	-3.5	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6