

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039201.D
 Acq On : 14 Dec 2023 12:07
 Operator : JC/MD
 Sample : 05719-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-RW10A-MW01S-GW-20231205

Quant Time: Dec 15 02:58:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214717	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	114734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92476	48.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.385	113	71274	48.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.647	98	272875	49.550	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	126351	55.608	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.220%
Target Compounds						
					Qvalue	
22) Diisopropyl ether	3.757	45	6865	1.384	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039201.D
Acq On : 14 Dec 2023 12:07
Operator : JC/MD
Sample : 05719-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-RW10A-MW01S-GW-20231205

Quant Time: Dec 15 02:58:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

