

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039205.D
 Acq On : 14 Dec 2023 14:02
 Operator : JC/MD
 Sample : 05829-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK-1

Quant Time: Dec 15 03:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	226549	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	117980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95936	48.413	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.820%
35) Dibromofluoromethane	5.385	113	74062	47.321	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	8.647	98	288211	49.602	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.079	95	131062	54.668	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.340%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	3411	2.065	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039205.D
Acq On : 14 Dec 2023 14:02
Operator : JC/MD
Sample : 05829-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK-1

Quant Time: Dec 15 03:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

