

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039207.D
 Acq On : 14 Dec 2023 14:48
 Operator : JC/MD
 Sample : 05829-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Dec 15 03:01:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	106389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219078	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94472	48.473	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.940%
35) Dibromofluoromethane	5.385	113	73647	48.660	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.320%
50) Toluene-d8	8.647	98	281234	50.051	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.079	95	129261	55.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.520%
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	2432	0.853	ug/l	94
52) Toluene	8.726	92	1281	0.314	ug/l	94
64) Tetrachloroethene	9.275	164	388	0.258	ug/l #	77
67) Ethyl Benzene	10.195	91	8386	0.899	ug/l	96
68) m/p-Xylenes	10.305	106	5606	1.675	ug/l	95
69) o-Xylene	10.646	106	1791	0.554	ug/l	89
80) 1,3,5-Trimethylbenzene	11.451	105	1730	0.226	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	8958	1.142	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039207.D
Acq On : 14 Dec 2023 14:48
Operator : JC/MD
Sample : 05829-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Dec 15 03:01:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

