

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039208.D
 Acq On : 14 Dec 2023 15:11
 Operator : JC/MD
 Sample : 05829-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Dec 15 03:01:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	227171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	236218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98367	48.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.385	113	75448	48.074	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	8.647	98	288597	49.532	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.060%
62) 4-Bromofluorobenzene	11.079	95	132049	54.929	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.860%
Target Compounds						
					Qvalue	
30) Chloroform	5.099	83	2432	0.816	ug/l	84
64) Tetrachloroethene	9.275	164	628	0.407	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039208.D
Acq On : 14 Dec 2023 15:11
Operator : JC/MD
Sample : 05829-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Dec 15 03:01:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

