

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039210.D
 Acq On : 14 Dec 2023 15:57
 Operator : JC/MD
 Sample : 05829-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Dec 15 03:02:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219697	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	113932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93917	48.417	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.840%
35) Dibromofluoromethane	5.385	113	72334	47.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	8.647	98	277696	49.283	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	127222	54.722	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.440%
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	82825	29.200	ug/l	98
64) Tetrachloroethene	9.268	164	894	0.601	ug/l	89
84) 1,2,4-Trimethylbenzene	11.756	105	2577	0.336	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039210.D
Acq On : 14 Dec 2023 15:57
Operator : JC/MD
Sample : 05829-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Dec 15 03:02:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

