

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
 Data File : VX039211.D
 Acq On : 14 Dec 2023 16:19
 Operator : JC/MD
 Sample : 05829-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Dec 15 03:02:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223497	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	118745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96888	48.656	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.320%
35) Dibromofluoromethane	5.379	113	74631	48.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.680%
50) Toluene-d8	8.647	98	286424	49.967	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	130827	55.316	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.640%
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	26827	9.213	ug/l	100
64) Tetrachloroethene	9.281	164	1698	1.099	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121423\
Data File : VX039211.D
Acq On : 14 Dec 2023 16:19
Operator : JC/MD
Sample : 05829-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Dec 15 03:02:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

