

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121520\
 Data File : VX020030.D
 Acq On : 15 Dec 2020 22:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 00:29:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	50.000	46.461	7.1	64	0.00
3 P	Chloromethane	50.000	43.789	12.4	66	0.00
4 C	Vinyl Chloride	50.000	50.118	-0.2#	71	0.00
5 T	Bromomethane	50.000	32.183	35.6#	46	0.00
6 T	Chloroethane	50.000	56.011	-12.0	78	0.00
7 T	Trichlorofluoromethane	50.000	51.188	-2.4	74	0.00
8 T	Diethyl Ether	50.000	52.395	-4.8	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.737	-7.5	77	0.00
10 T	Methyl Iodide	50.000	29.422	41.2#	42	0.00
11 T	Tert butyl alcohol	250.000	274.877	-10.0	82	0.00
12 CM	1,1-Dichloroethene	50.000	52.689	-5.4#	76	0.00
13 T	Acrolein	250.000	248.172	0.7	76	0.00
14 T	Allyl chloride	50.000	53.579	-7.2	78	0.00
15 T	Acrylonitrile	250.000	294.141	-17.7	87	0.00
16 T	Acetone	250.000	275.621	-10.2	80	0.00
17 T	Carbon Disulfide	50.000	43.369	13.3	64	0.00
18 T	Methyl Acetate	50.000	63.879	-27.8#	92	0.00
19 T	Methyl tert-butyl Ether	50.000	56.223	-12.4	81	0.00
20 T	Methylene Chloride	50.000	54.740	-9.5	77	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.487	-3.0	74	0.00
22 T	Diisopropyl ether	50.000	58.273	-16.5	84	0.00
23 T	Vinyl Acetate	250.000	290.230	-16.1	82	0.00
24 P	1,1-Dichloroethane	50.000	55.814	-11.6	81	0.00
25 T	2-Butanone	250.000	292.049	-16.8	85	0.00
26 T	2,2-Dichloropropane	50.000	47.103	5.8	68	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.239	-6.5	78	0.00
28 T	Bromochloromethane	50.000	53.332	-6.7	78	0.00
29 T	Tetrahydrofuran	250.000	289.373	-15.7	85	0.00
30 C	Chloroform	50.000	56.409	-12.8#	82	0.00
31 T	Cyclohexane	50.000	50.900	-1.8	72	0.00
32 T	1,1,1-Trichloroethane	50.000	54.594	-9.2	78	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.887	0.2	77	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	76	0.00
35 S	Dibromofluoromethane	50.000	48.692	2.6	76	0.00
36 T	1,1-Dichloropropene	50.000	49.157	1.7	75	0.00
37 T	Ethyl Acetate	50.000	54.493	-9.0	85	0.00
38 T	Carbon Tetrachloride	50.000	51.016	-2.0	75	0.00
39 T	Methylcyclohexane	50.000	47.620	4.8	70	0.00
40 TM	Benzene	50.000	51.587	-3.2	78	0.00
41 T	Methacrylonitrile	50.000	54.395	-8.8	83	0.00
42 TM	1,2-Dichloroethane	50.000	52.605	-5.2	79	0.00
43 T	Isopropyl Acetate	50.000	53.717	-7.4	83	0.00
44 TM	Trichloroethene	50.000	48.608	2.8	73	0.00
45 C	1,2-Dichloropropane	50.000	54.328	-8.7#	82	0.00
46 T	Dibromomethane	50.000	52.793	-5.6	79	0.00
47 T	Bromodichloromethane	50.000	54.500	-9.0	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121520\
 Data File : VX020030.D
 Acq On : 15 Dec 2020 22:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 00:29:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	54.585	-9.2	83	0.00
49 T	1,4-Dioxane	1000.000	1056.807	-5.7	80	0.00
50 S	Toluene-d8	50.000	47.913	4.2	74	0.00
51 T	4-Methyl-2-Pentanone	250.000	286.003	-14.4	87	0.00
52 CM	Toluene	50.000	51.434	-2.9#	77	0.00
53 T	t-1,3-Dichloropropene	50.000	51.502	-3.0	75	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.942	-3.9	76	0.00
55 T	1,1,2-Trichloroethane	50.000	53.995	-8.0	81	0.00
56 T	Ethyl methacrylate	50.000	54.926	-9.9	81	0.00
57 T	1,3-Dichloropropane	50.000	53.564	-7.1	82	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	276.133	-10.5	83	0.00
59 T	2-Hexanone	250.000	281.036	-12.4	84	0.00
60 T	Dibromochloromethane	50.000	54.473	-8.9	80	0.00
61 T	1,2-Dibromoethane	50.000	52.498	-5.0	80	0.00
62 S	4-Bromofluorobenzene	50.000	49.071	1.9	76	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	76	0.00
64 T	Tetrachloroethene	50.000	47.408	5.2	71	0.00
65 PM	Chlorobenzene	50.000	50.523	-1.0	76	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.050	-8.1	80	0.00
67 C	Ethyl Benzene	50.000	52.132	-4.3#	77	0.00
68 T	m/p-Xylenes	100.000	102.614	-2.6	76	0.00
69 T	o-Xylene	50.000	52.190	-4.4	77	0.00
70 T	Styrene	50.000	53.698	-7.4	77	0.00
71 P	Bromoform	50.000	54.086	-8.2	77	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	73	0.00
73 T	Isopropylbenzene	50.000	53.115	-6.2	78	0.00
74 T	N-amyl acetate	50.000	55.433	-10.9	80	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	55.871	-11.7	83	0.00
76 T	1,2,3-Trichloropropane	50.000	53.672	-7.3	79	0.00
77 T	Bromobenzene	50.000	51.274	-2.5	77	0.00
78 T	n-propylbenzene	50.000	52.505	-5.0	76	0.00
79 T	2-Chlorotoluene	50.000	52.497	-5.0	77	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.787	-5.6	75	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.358	-4.7	74	0.00
82 T	4-Chlorotoluene	50.000	52.616	-5.2	78	0.00
83 T	tert-Butylbenzene	50.000	53.286	-6.6	77	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.016	-8.0	76	0.00
85 T	sec-Butylbenzene	50.000	53.862	-7.7	77	0.00
86 T	p-Isopropyltoluene	50.000	53.827	-7.7	76	0.00
87 T	1,3-Dichlorobenzene	50.000	50.192	-0.4	73	0.00
88 T	1,4-Dichlorobenzene	50.000	50.439	-0.9	75	0.00
89 T	n-Butylbenzene	50.000	52.703	-5.4	74	0.00
90 T	Hexachloroethane	50.000	54.449	-8.9	79	0.00
91 T	1,2-Dichlorobenzene	50.000	50.896	-1.8	74	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	56.668	-13.3	84	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.278	-4.6	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121520\
Data File : VX020030.D
Acq On : 15 Dec 2020 22:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 16 00:29:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	51.572	-3.1	78	0.00
95 T	Naphthalene	50.000	56.361	-12.7	80	0.00
96 T	1,2,3-Trichlorobenzene	50.000	54.455	-8.9	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6