

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121520\
 Data File : VX020032.D
 Acq On : 15 Dec 2020 23:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 01:37:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.458	0.421	8.1	65	0.00
3 P	Chloromethane	0.538	0.448	16.7	65	0.00
4 C	Vinyl Chloride	0.568	0.559	1.6#	72	0.00
5 T	Bromomethane	0.188	0.139	26.1#	55	0.00
6 T	Chloroethane	0.341	0.368	-7.9	77	0.00
7 T	Trichlorofluoromethane	0.791	0.800	-1.1	75	0.00
8 T	Diethyl Ether	0.335	0.348	-3.9	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.475	-4.9	77	0.00
10 T	Methyl Iodide	0.668	0.444	33.5#	49#	0.00
11 T	Tert butyl alcohol	0.145	0.148	-2.1	84	0.00
12 CM	1,1-Dichloroethene	0.458	0.466	-1.7#	75	0.00
13 T	Acrolein	0.073	0.058	20.5	63	0.00
14 T	Allyl chloride	0.774	0.805	-4.0	78	0.00
15 T	Acrylonitrile	0.297	0.338	-13.8	86	0.00
16 T	Acetone	0.247	0.268	-8.5	81	0.00
17 T	Carbon Disulfide	1.314	1.111	15.4	64	0.00
18 T	Methyl Acetate	0.604	0.745	-23.3	91	0.00
19 T	Methyl tert-butyl Ether	1.668	1.838	-10.2	81	0.00
20 T	Methylene Chloride	0.639	0.601	5.9	78	0.00
21 T	trans-1,2-Dichloroethene	0.534	0.545	-2.1	75	0.00
22 T	Diisopropyl ether	1.564	1.788	-14.3	84	0.00
23 T	Vinyl Acetate	1.333	1.510	-13.3	82	0.00
24 P	1,1-Dichloroethane	0.943	1.031	-9.3	82	0.00
25 T	2-Butanone	0.380	0.429	-12.9	85	0.00
26 T	2,2-Dichloropropane	0.812	0.746	8.1	68	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.650	-4.0	79	0.00
28 T	Bromochloromethane	0.450	0.476	-5.8	79	0.00
29 T	Tetrahydrofuran	0.245	0.276	-12.7	85	0.00
30 C	Chloroform	0.970	1.068	-10.1#	82	0.00
31 T	Cyclohexane	0.817	0.814	0.4	72	0.00
32 T	1,1,1-Trichloroethane	0.835	0.892	-6.8	78	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.658	-0.2	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.323	0.319	1.2	78	0.00
36 T	1,1-Dichloropropene	0.452	0.442	2.2	75	0.00
37 T	Ethyl Acetate	0.456	0.487	-6.8	84	0.00
38 T	Carbon Tetrachloride	0.442	0.446	-0.9	75	0.00
39 T	Methylcyclohexane	0.531	0.497	6.4	70	0.00
40 TM	Benzene	1.380	1.418	-2.8	79	0.00
41 T	Methacrylonitrile	0.250	0.271	-8.4	83	0.00
42 TM	1,2-Dichloroethane	0.472	0.497	-5.3	80	0.00
43 T	Isopropyl Acetate	0.742	0.787	-6.1	82	0.00
44 TM	Trichloroethene	0.362	0.359	0.8	76	0.00
45 C	1,2-Dichloropropane	0.345	0.373	-8.1#	83	0.00
46 T	Dibromomethane	0.233	0.248	-6.4	81	0.00
47 T	Bromodichloromethane	0.462	0.497	-7.6	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121520\
 Data File : VX020032.D
 Acq On : 15 Dec 2020 23:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 01:37:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	0.365	0.394	-7.9	83	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	82	0.00
50 S	Toluene-d8	1.234	1.221	1.1	77	0.00
51 T	4-Methyl-2-Pentanone	0.453	0.509	-12.4	86	0.00
52 CM	Toluene	0.863	0.887	-2.8#	78	0.00
53 T	t-1,3-Dichloropropene	0.513	0.529	-3.1	76	0.00
54 T	cis-1,3-Dichloropropene	0.557	0.581	-4.3	77	0.00
55 T	1,1,2-Trichloroethane	0.348	0.379	-8.9	83	0.00
56 T	Ethyl methacrylate	0.526	0.572	-8.7	81	0.00
57 T	1,3-Dichloropropane	0.594	0.635	-6.9	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.279	0.310	-11.1	85	0.00
59 T	2-Hexanone	0.351	0.388	-10.5	84	0.00
60 T	Dibromochloromethane	0.353	0.378	-7.1	79	0.00
61 T	1,2-Dibromoethane	0.369	0.385	-4.3	80	0.00
62 S	4-Bromofluorobenzene	0.470	0.472	-0.4	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.335	0.333	0.6	76	0.00
65 PM	Chlorobenzene	0.996	1.002	-0.6	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.382	-6.4	81	0.00
67 C	Ethyl Benzene	1.768	1.804	-2.0#	77	0.00
68 T	m/p-Xylenes	0.666	0.675	-1.4	76	0.00
69 T	o-Xylene	0.641	0.651	-1.6	76	0.00
70 T	Styrene	1.082	1.142	-5.5	78	0.00
71 P	Bromoform	0.282	0.301	-6.7	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.416	3.604	-5.5	78	0.00
74 T	N-amyl acetate	1.369	1.485	-8.5	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.204	1.307	-8.6	81	0.00
76 T	1,2,3-Trichloropropane	1.121	1.222	-9.0	81	0.00
77 T	Bromobenzene	0.874	0.912	-4.3	79	0.00
78 T	n-propylbenzene	3.902	4.079	-4.5	76	0.00
79 T	2-Chlorotoluene	2.403	2.482	-3.3	77	0.00
80 T	1,3,5-Trimethylbenzene	2.883	2.999	-4.0	75	0.00
81 T	trans-1,4-Dichloro-2-butene	0.388	0.392	-1.0	73	0.00
82 T	4-Chlorotoluene	2.814	2.925	-3.9	78	0.00
83 T	tert-Butylbenzene	2.828	2.966	-4.9	76	0.00
84 T	1,2,4-Trimethylbenzene	2.904	3.106	-7.0	77	0.00
85 T	sec-Butylbenzene	3.228	3.458	-7.1	77	0.00
86 T	p-Isopropyltoluene	2.959	3.121	-5.5	75	0.00
87 T	1,3-Dichlorobenzene	1.604	1.550	3.4	72	0.00
88 T	1,4-Dichlorobenzene	1.634	1.594	2.4	73	0.00
89 T	n-Butylbenzene	2.621	2.703	-3.1	74	0.00
90 T	Hexachloroethane	0.511	0.542	-6.1	78	0.00
91 T	1,2-Dichlorobenzene	1.564	1.636	-4.6	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.259	0.282	-8.9	82	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.074	-5.5	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121520\
Data File : VX020032.D
Acq On : 15 Dec 2020 23:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 16 01:37:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	0.455	0.474	-4.2	80	0.00
95 T	Naphthalene	3.301	3.717	-12.6	81	0.00
96 T	1,2,3-Trichlorobenzene	0.995	1.065	-7.0	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6