

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121521\
 Data File : VX025806.D
 Acq On : 15 Dec 2021 10:44
 Operator : JC/MD
 Sample : VX1215WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK653

Quant Time: Dec 16 00:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 16 00:37:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	184730	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	177365	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	81281	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	47270	37.495	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.000%
7) Chloroethane-d5	1.666	69	43697	41.996	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.307	63	71692	35.665	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.340%
21) 2-Butanone-d5	4.453	46	93070	99.173	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.170%
24) Chloroform-d	5.062	84	109293	49.604	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.958	65	73291	53.366	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.740%
32) Benzene-d6	5.977	84	212208	45.144	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.280%
36) 1,2-Dichloropropane-d6	7.306	67	65074	44.712	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.420%
41) Toluene-d8	8.647	98	190604	42.958	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.920%
43) trans-1,3-Dichloroprop...	8.952	79	26427	36.116	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.240%
47) 2-Hexanone-d5	9.385	63	64675	85.369	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.370%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	96711	47.018	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.040%
66) 1,2-Dichlorobenzene-d4	12.323	152	72596	45.538	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121521\
Data File : VX025806.D
Acq On : 15 Dec 2021 10:44
Operator : JC/MD
Sample : VX1215WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK653

Quant Time: Dec 16 00:37:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM112221WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 16 00:37:00 2021
Response via : Initial Calibration

