

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033431.D
 Acq On : 15 Dec 2022 10:39
 Operator : JC/MD
 Sample : VX1215WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1215WBL01

Quant Time: Dec 16 00:59:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223781	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92099	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39540	42.991	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.980%
35) Dibromofluoromethane	5.385	113	44225	43.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.880%
50) Toluene-d8	8.647	98	90233	42.854	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	85.700%#
62) 4-Bromofluorobenzene	11.079	95	70078	43.528	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033431.D
Acq On : 15 Dec 2022 10:39
Operator : JC/MD
Sample : VX1215WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1215WBL01

Quant Time: Dec 16 00:59:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

