

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033435.D
 Acq On : 15 Dec 2022 12:32
 Operator : JC/MD
 Sample : N6070-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-226-245-121422

Quant Time: Dec 16 01:01:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	87583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37277	66.344	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 132.680%#			
35) Dibromofluoromethane	5.385	113	30503	36.410	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 72.820%#			
50) Toluene-d8	8.647	98	86490	49.351	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.700%			
62) 4-Bromofluorobenzene	11.079	95	65521	48.895	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.800%			
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	1944	2.448	ug/l #	75
16) Acetone	2.380	43	1934	4.652	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	9131	8.796	ug/l	86
44) Trichloroethene	7.129	130	6654	5.068	ug/l	87
52) Toluene	8.720	92	772896	251.014	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033435.D
Acq On : 15 Dec 2022 12:32
Operator : JC/MD
Sample : N6070-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-226-245-121422

Quant Time: Dec 16 01:01:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

