

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033437.D
 Acq On : 15 Dec 2022 13:18
 Operator : JC/MD
 Sample : N6070-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-02-160-180-121422

Quant Time: Dec 16 01:01:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176885	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	31663	44.351	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.700%
35) Dibromofluoromethane	5.385	113	35505	44.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.260%
50) Toluene-d8	8.653	98	72776	43.727	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.460%#
62) 4-Bromofluorobenzene	11.079	95	53495	42.037	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.080%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	1234	2.336	ug/l #	83
30) Chloroform	5.098	83	6355	2.822	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033437.D
Acq On : 15 Dec 2022 13:18
Operator : JC/MD
Sample : N6070-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-02-160-180-121422

Quant Time: Dec 16 01:01:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

