

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033439.D
 Acq On : 15 Dec 2022 14:04
 Operator : JC/MD
 Sample : PB149669TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149669TB

Quant Time: Dec 16 01:02:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38125	50.414	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.820%	
35) Dibromofluoromethane	5.385	113	42223	49.743	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.480%	
50) Toluene-d8	8.647	98	84528	47.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.200%	
62) 4-Bromofluorobenzene	11.079	95	64652	47.618	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.240%	
Target Compounds						
20) Methylene Chloride	2.788	84	5621	4.504	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033439.D
Acq On : 15 Dec 2022 14:04
Operator : JC/MD
Sample : PB149669TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149669TB

Quant Time: Dec 16 01:02:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

