

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033440.D
 Acq On : 15 Dec 2022 14:27
 Operator : JC/MD
 Sample : N6051-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-591

Quant Time: Dec 16 01:02:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	120878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37423	48.258	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.385	113	41143	48.228	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	8.646	98	84379	47.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.079	95	64564	47.315	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.620%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	16950	29.538	ug/l	94
20) Methylene Chloride	2.788	84	164143	128.263	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033440.D
Acq On : 15 Dec 2022 14:27
Operator : JC/MD
Sample : N6051-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-591

Quant Time: Dec 16 01:02:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

