

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033441.D
 Acq On : 15 Dec 2022 14:50
 Operator : JC/MD
 Sample : N6070-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR-04-226-245-121422

Quant Time: Dec 16 01:02:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36088	48.816	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.640%		
35) Dibromofluoromethane	5.385	113	40491	49.290	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.580%		
50) Toluene-d8	8.653	98	81865	47.637	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.280%		
62) 4-Bromofluorobenzene	11.079	95	63928	48.651	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.300%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	1845	1.766	ug/l	83
16) Acetone	2.373	43	2007	3.669	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	12282	8.993	ug/l	87
44) Trichloroethene	7.129	130	6252	4.856	ug/l	95
52) Toluene	8.720	92	739140	244.804	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033441.D
Acq On : 15 Dec 2022 14:50
Operator : JC/MD
Sample : N6070-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR-04-226-245-121422

Quant Time: Dec 16 01:02:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

