

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033442.D
 Acq On : 15 Dec 2022 15:13
 Operator : JC/MD
 Sample : PB149652TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149652TB

Quant Time: Dec 16 01:03:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	135318	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	24590	45.399	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	30909	50.784	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.560%
50) Toluene-d8	8.647	98	64277	50.483	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	46774	48.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11012	27.475	ug/l	90
20) Methylene Chloride	2.794	84	3351	3.749	ug/l #	83
37) Ethyl Acetate	4.721	43	11586	10.329	ug/l	98
43) Isopropyl Acetate	6.336	43	50974	28.243	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033442.D
Acq On : 15 Dec 2022 15:13
Operator : JC/MD
Sample : PB149652TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149652TB

Quant Time: Dec 16 01:03:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

