

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033445.D
 Acq On : 15 Dec 2022 16:22
 Operator : JC/MD
 Sample : N6065-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-S

Quant Time: Dec 16 01:04:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190891	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	39650	53.775	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.540%	
35) Dibromofluoromethane	5.385	113	43747	50.952	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.900%	
50) Toluene-d8	8.647	98	90574	50.427	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.860%	
62) 4-Bromofluorobenzene	11.079	95	69609	50.686	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.380%	
Target Compounds						
16) Acetone	2.380	43	15968	29.266	ug/l	95
20) Methylene Chloride	2.794	84	5192	4.267	ug/l	86
37) Ethyl Acetate	4.714	43	15646	9.888	ug/l	99
43) Isopropyl Acetate	6.342	43	71269	27.992	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033445.D
Acq On : 15 Dec 2022 16:22
Operator : JC/MD
Sample : N6065-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-S

Quant Time: Dec 16 01:04:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

