

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
 Data File : VX033446.D
 Acq On : 15 Dec 2022 16:45
 Operator : JC/MD
 Sample : N6067-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-SOIL-20221214

Quant Time: Dec 16 01:04:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38829	51.791	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.385	113	42408	49.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%
50) Toluene-d8	8.647	98	88562	49.678	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	68474	50.235	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.460%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12294	22.160	ug/l	94
20) Methylene Chloride	2.794	84	4365	3.528	ug/l	88
37) Ethyl Acetate	4.715	43	13830	8.806	ug/l	98
43) Isopropyl Acetate	6.336	43	65913	26.083	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121522\
Data File : VX033446.D
Acq On : 15 Dec 2022 16:45
Operator : JC/MD
Sample : N6067-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-SOIL-20221214

Quant Time: Dec 16 01:04:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

