

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121523\
 Data File : VX039228.D
 Acq On : 15 Dec 2023 12:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 16 01:14:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	10047	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	78941	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	95593	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	45691	33.242	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.800%#
60) 4-Bromofluorobenzene	11.079	95	64607	30.943	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	103.133%
63) Toluene-d8	8.647	98	114660	30.338	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121523\
Data File : VX039228.D
Acq On : 15 Dec 2023 12:49
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 16 01:14:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

