

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
 Data File : VX048872.D
 Acq On : 15 Dec 2025 12:10
 Operator : JC/MD
 Sample : Q3835-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-01-170-121025

Quant Time: Dec 16 00:14:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	202756	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	340422	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	306263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	138818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	132925	48.845	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.700%
35) Dibromofluoromethane	5.367	113	113696	48.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	8.628	98	407388	49.583	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.061	95	139407	49.205	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.110	130	3836	1.545	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
Data File : VX048872.D
Acq On : 15 Dec 2025 12:10
Operator : JC/MD
Sample : Q3835-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-01-170-121025

Quant Time: Dec 16 00:14:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

