

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
 Data File : VX048877.D
 Acq On : 15 Dec 2025 14:44
 Operator : JC/MD
 Sample : Q3848-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-57.5-121125-FDDL

Quant Time: Dec 16 00:16:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	197001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	326976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	273805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	100588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	137015	51.819	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.640%
35) Dibromofluoromethane	5.367	113	111756	49.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.280%
50) Toluene-d8	8.629	98	387995	49.165	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.320%
62) 4-Bromofluorobenzene	11.061	95	107849	39.631	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.260%#
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	11102	4.116	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.337	101	1313	0.586	ug/l	97
27) cis-1,2-Dichloroethene	4.471	96	170219	64.608	ug/l	94
44) Trichloroethene	7.104	130	72003	30.188	ug/l	98
64) Tetrachloroethene	9.250	164	947	0.474	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
Data File : VX048877.D
Acq On : 15 Dec 2025 14:44
Operator : JC/MD
Sample : Q3848-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-57.5-121125-FDDL

Quant Time: Dec 16 00:16:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

