

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
 Data File : VX048878.D
 Acq On : 15 Dec 2025 15:08
 Operator : JC/MD
 Sample : Q3863-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17B-55-121225DL

Quant Time: Dec 16 00:17:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	182137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	301665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	265370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	111258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	126601	51.788	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.367	113	103147	49.660	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.320%
50) Toluene-d8	8.628	98	363041	49.863	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.061	95	112630	44.861	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.720%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.386	62	713	0.286	ug/l #	84
9) 1,1,2-Trichlorotrifluo...	2.337	101	2745	1.325	ug/l	97
27) cis-1,2-Dichloroethene	4.471	96	66900	27.465	ug/l	96
44) Trichloroethene	7.104	130	149597	67.983	ug/l	98
64) Tetrachloroethene	9.262	164	1992	1.030	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121525\
Data File : VX048878.D
Acq On : 15 Dec 2025 15:08
Operator : JC/MD
Sample : Q3863-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17B-55-121225DL

Quant Time: Dec 16 00:17:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

