

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121619\
 Data File : VX014036.D
 Acq On : 16 Dec 2019 15:37
 Operator : JC/SP
 Sample : K6252-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20191210

Quant Time: Dec 17 04:40:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	577650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	886771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	788331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	341237	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	321057	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	5.49	113	270133	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	8.71	98	1053133	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	11.14	95	348325	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	21776	3.988	ug/l	97
27) cis-1,2-Dichloroethene	4.58	96	22305	3.222	ug/l	84
44) Trichloroethene	7.20	130	81990	11.857	ug/l	99
45) 1,2-Dichloropropane	7.51	63	19761	3.080	ug/l	93
64) Tetrachloroethene	9.33	164	69375	9.945	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121619\
Data File : VX014036.D
Acq On : 16 Dec 2019 15:37
Operator : JC/SP
Sample : K6252-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20191210

Quant Time: Dec 17 04:40:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

