

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121619\
 Data File : VX014041.D
 Acq On : 16 Dec 2019 17:34
 Operator : JC/SP
 Sample : K6251-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D1-20191206

Quant Time: Dec 17 04:55:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	567538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	877393	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	780163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	327791	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	319823	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
35) Dibromofluoromethane	5.49	113	258502	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	1033640	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	11.14	95	341400	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
Target Compounds						
16) Acetone	2.43	43	13354	3.198	ug/l	92
27) cis-1,2-Dichloroethene	4.58	96	3548	0.522	ug/l	87
38) Carbon Tetrachloride	5.77	117	7270	0.991	ug/l #	85
44) Trichloroethene	7.21	130	391546	57.228	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	2296	0.366	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121619\
Data File : VX014041.D
Acq On : 16 Dec 2019 17:34
Operator : JC/SP
Sample : K6251-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D1-20191206

Quant Time: Dec 17 04:55:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

