

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121619\
 Data File : VX014043.D
 Acq On : 16 Dec 2019 18:20
 Operator : JC/SP
 Sample : K6251-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20191206

Quant Time: Dec 17 05:33:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	567781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	873510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	771740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	327373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	319410	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
35) Dibromofluoromethane	5.49	113	262160	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.71	98	1034178	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.13	95	336322	44.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.02%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	27348	5.095	ug/l	97
12) 1,1-Dichloroethene	2.37	96	3725	0.682	ug/l	90
16) Acetone	2.43	43	10929	2.616	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	8326	1.224	ug/l	79
44) Trichloroethene	7.20	130	548165	80.475	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121619\
Data File : VX014043.D
Acq On : 16 Dec 2019 18:20
Operator : JC/SP
Sample : K6251-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20191206

Quant Time: Dec 17 05:33:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

