

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX121619\
 Data File : VX014023.D
 Acq On : 16 Dec 2019 10:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

apatel
 12/17/2019 4:50:04 PM

Quant Time: Dec 17 03:37:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	650740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	977539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	877742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	439040	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.04	65	348498	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	
35) Dibromofluoromethane	5.48	113	284634	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
50) Toluene-d8	8.71	98	1110544	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
62) 4-Bromofluorobenzene	11.13	95	394862	46.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.38%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.19	85	279776	48.057	ug/l	97
3) Chloromethane	1.32	50	366140	46.864	ug/l	99
4) Vinyl Chloride	1.40	62	385164	46.721	ug/l	100
5) Bromomethane	1.63	94	249962	42.720	ug/l	99
6) Chloroethane	1.71	64	242528	47.666	ug/l	95
7) Trichlorofluoromethane	1.92	101	479571	46.993	ug/l	100
8) Diethyl Ether	2.18	74	211913	44.245	ug/l	97
9) 1,1,2-Trichlorotrifluoroet	2.37	101	294269	47.835	ug/l	99
10) Methyl Iodide	2.50	142	388459	47.994	ug/l	99
11) Tert butyl alcohol	3.03	59	324992	203.933	ug/l	99
12) 1,1-Dichloroethene	2.36	96	285682	45.638	ug/l	98
13) Acrolein	2.28	56	301595	330.610	ug/l	98
14) Allyl chloride	2.72	41	543276	48.608	ug/l	100
15) Acrylonitrile	3.13	53	845220	227.233	ug/l	99
16) Acetone	2.43	43	1196744	249.941	ug/l	100
17) Carbon Disulfide	2.56	76	826279	47.584	ug/l	100
18) Methyl Acetate	2.76	43	426114	44.912	ug/l	99
19) Methyl tert-butyl Ether	3.18	73	966613	46.992	ug/l	98
20) Methylene Chloride	2.84	84	331557	43.973	ug/l	97
21) trans-1,2-Dichloroethene	3.15	96	312681	45.341	ug/l	98
22) Diisopropyl ether	3.84	45	1053210	47.289	ug/l	97
23) Vinyl Acetate	3.80	43	4493937	247.256	ug/l	100
24) 1,1-Dichloroethane	3.69	63	573601	46.291	ug/l	99
25) 2-Butanone	4.65	43	1439048	243.775	ug/l	100
26) 2,2-Dichloropropane	4.56	77	474716	49.343	ug/l	100
27) cis-1,2-Dichloroethene	4.58	96	360754	46.255	ug/l	100
28) Bromochloromethane	5.00	49	248095	52.042	ug/l	97
29) Tetrahydrofuran	5.11	42	758976	229.867	ug/l	98
30) Chloroform	5.20	83	550649	46.877	ug/l	96
31) Cyclohexane	5.57	56	531680	48.243	ug/l	97
32) 1,1,1-Trichloroethane	5.48	97	469248	46.872	ug/l	99
36) 1,1-Dichloropropene	5.79	75	430143	47.956	ug/l	99
37) Ethyl Acetate	4.81	43	465070	46.954	ug/l	100
38) Carbon Tetrachloride	5.77	117	404296	49.479	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121619\
 Data File : VX014023.D
 Acq On : 16 Dec 2019 10:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050

Manual Integrations
 APPROVED

apatel
 12/17/2019 4:50:04 PM

Quant Time: Dec 17 03:37:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.45	83	546743	49.367	ug/l	98
40) Benzene	6.13	78	1295678	46.948	ug/l	100
41) Methacrylonitrile	5.02	41	260216	47.349	ug/l	98
42) 1,2-Dichloroethane	6.18	62	445130	47.582	ug/l	98
43) Isopropyl Acetate	6.43	43	774355	47.240	ug/l	100
44) Trichloroethene	7.20	130	347833	45.631	ug/l	100
45) 1,2-Dichloropropane	7.50	63	338061	47.792	ug/l	98
46) Dibromomethane	7.65	93	217459	46.498	ug/l	98
47) Bromodichloromethane	7.89	83	438682	49.805	ug/l	100
48) Methyl methacrylate	7.76	41	385076	47.974	ug/l	99
49) 1,4-Dioxane	7.73	88	142282	871.049	ug/l	99
51) 4-Methyl-2-Pentanone	8.64	43	2366328	229.864	ug/l	100
52) Toluene	8.78	92	811227	46.505	ug/l	99
53) t-1,3-Dichloropropene	9.03	75	492403	50.984	ug/l	98
54) cis-1,3-Dichloropropene	8.43	75	549690	51.088	ug/l	100
55) 1,1,2-Trichloroethane	9.21	97	323112	46.216	ug/l	99
56) Ethyl methacrylate	9.17	69	530780	48.331	ug/l	99
57) 1,3-Dichloropropane	9.36	76	548890	46.663	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.30	63	1100864	264.882	ug/l	99
59) 2-Hexanone	9.48	43	2021919	243.985	ug/l	98
60) Dibromochloromethane	9.57	129	353367	50.848	ug/l	100
61) 1,2-Dibromoethane	9.67	107	343988	48.089	ug/l	99
64) Tetrachloroethene	9.33	164	336399	43.312	ug/l	99
65) Chlorobenzene	10.14	112	865726	46.388	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.21	131	320595	48.110	ug/l	99
67) Ethyl Benzene	10.25	91	1556159	48.500	ug/l	100
68) m/p-Xylenes	10.35	106	1185875	96.095	ug/l	100
69) o-Xylene	10.70	106	569706	47.071	ug/l	97
70) Styrene	10.71	104	989969	48.371	ug/l	98
71) Bromoform	10.85	173	269651	50.341	ug/l	100
73) Isopropylbenzene	11.01	105	1511830	48.908	ug/l	99
74) N-amyl acetate	10.89	43	708291	48.883	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.26	83	504024	47.031	ug/l	100
76) 1,2,3-Trichloropropane	11.29	75	403054m	42.287	ug/l	
77) Bromobenzene	11.25	156	392452	46.144	ug/l	99
78) n-propylbenzene	11.35	91	1745388	50.339	ug/l	100
79) 2-Chlorotoluene	11.42	91	1019503	48.371	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	1270462	48.835	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	175394	51.939	ug/l	98
82) 4-Chlorotoluene	11.51	91	1182296	48.359	ug/l	99
83) tert-Butylbenzene	11.76	119	1258854	49.723	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	1281327	49.181	ug/l	100
85) sec-Butylbenzene	11.94	105	1486682	49.742	ug/l	99
86) p-Isopropyltoluene	12.06	119	1378464	50.421	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	692642	46.468	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	692467	45.688	ug/l	100
89) n-Butylbenzene	12.39	91	1203074	51.572	ug/l	100
90) Hexachloroethane	12.59	117	251038	51.121	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	694225	46.266	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	12.99	75	104938	44.248	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121619\
Data File : VX014023.D
Acq On : 16 Dec 2019 10:02
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM

Quant Time: Dec 17 03:37:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.64	180	479619	49.178	ug/l	98
94) Hexachlorobutadiene	13.78	225	222154	47.397	ug/l	97
95) Naphthalene	13.83	128	1428881	49.867	ug/l	100
96) 1,2,3-Trichlorobenzene	14.01	180	466337	48.439	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1





Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 5.65 min Scan# 781
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
168	100		
99	47.7	40.3	60.5

Manual Integrations
 APPROVED




apatel
 12/17/2019 4:50:04 PM




#2

Dichlorodifluoromethane
 Concen: 48.057 ug/l
 RT: 1.19 min Scan# 50
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
85	100		
87	31.0	16.4	49.2





#3

Chloromethane

Concen: 46.864 ug/l

RT: 1.32 min Scan# 71

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 50 Resp: 366140

Ion Ratio Lower Upper

50 100





52 33.3 26.2 39.4

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#4

Vinyl Chloride

Concen: 46.721 ug/l

RT: 1.40 min Scan# 84

Delta R.T. 0.00 min

Lab File: VX014023.D




Acq: 16 Dec 2019 10:02


Tgt Ion: 62 Resp: 385164

Ion Ratio Lower Upper

62 100

64 32.1 25.7 38.5





#5

Bromomethane

Concen: 42.720 ug/l

RT: 1.63 min Scan# 122

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 94 Resp: 249962

Ion Ratio Lower Upper

94 100





96 95.4 75.2 112.8

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#6

Chloroethane

Concen: 47.666 ug/l

RT: 1.71 min Scan# 135

Delta R.T. 0.01 min

Lab File: VX014023.D




Acq: 16 Dec 2019 10:02


Tgt Ion: 64 Resp: 242528

Ion Ratio Lower Upper

64 100

66 32.2 23.4 35.2





#7





Trichlorofluoromethane
Concen: 46.993 ug/l
RT: 1.92 min Scan# 169
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ion	Ratio	Resp	Lower	Upper
101	100		479571		
103	65.1	52.2	78.4		

Manual Integrations
APPROVED




apatel
12/17/2019 4:50:04 PM




#8

Diethyl Ether
Concen: 44.245 ug/l
RT: 2.18 min Scan# 212
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
74	100		211913		
45	99.3	48.1	144.3		





#9

1,1,2-Trichlorotrifluoroethane

Concen: 47.835 ug/l

RT: 2.37 min Scan# 243

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	101	Resp:	294269
Ion Ratio	Lower	Upper	
101	100		
85	41.9	33.7	50.5
151	79.4	64.5	96.7

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#10

Methyl Iodide

Concen: 47.994 ug/l




RT: 2.50 min Scan# 265


Delta R.T. 0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	142	Resp:	388459
Ion Ratio	Lower	Upper	
142	100		
127	40.2	31.6	47.4
141	14.2	11.6	17.4





#11





Tert butyl alcohol
Concen: 203.933 ug/l
RT: 3.03 min Scan# 352
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
59	100		
57	10.2	8.4	12.6

Manual Integrations
APPROVED




apatel
12/17/2019 4:50:04 PM




#12

1,1-Dichloroethene
Concen: 45.638 ug/l
RT: 2.36 min Scan# 242
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
96	100		
61	162.7	127.9	191.9
98	65.8	50.5	75.7





#13

Acrolein

Concen: 330.610 ug/l

RT: 2.28 min Scan# 229

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 56 Resp: 301595

Ion Ratio Lower Upper

56 100





55 69.8 56.9 85.3

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#14

Allyl chloride

Concen: 48.608 ug/l

RT: 2.72 min Scan# 300

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02




Tgt Ion: 41 Resp: 543276


Ion Ratio Lower Upper

41 100

39 64.8 51.8 77.8

76 32.5 25.9 38.9





#15

Acrylonitrile

Concen: 227.233 ug/l

RT: 3.13 min Scan# 368

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	53	Resp:	845220
Ion Ratio		Lower	Upper
53	100		
52	83.5	66.5	99.7
51	35.9	28.1	42.1

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#16

Acetone

Concen: 249.941 ug/l




RT: 2.43 min Scan# 253


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	43	Resp:	1196744
Ion Ratio		Lower	Upper
43	100		
58	30.8	24.9	37.3





#17

Carbon Disulfide

Concen: 47.584 ug/l

RT: 2.56 min Scan# 274

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 76 Resp: 826279

Ion Ratio Lower Upper

76 100


78 9.1 7.2 10.8

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.56

500000

400000



300000

200000

100000

0

Time-->



#18

Methyl Acetate

Concen: 44.912 ug/l

RT: 2.76 min Scan# 307

Delta R.T. -0.01 min

Lab File: VX014023.D


Acq: 16 Dec 2019 10:02

Tgt Ion: 43 Resp: 426114

Ion Ratio Lower Upper

43 100

74 23.8 19.5 29.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

2.76

250000

200000


150000


100000

50000

0

Time-->









#19
Methvl tert-butvl Ether
Concen: 46.992 ug/l
RT: 3.18 min Scan# 376
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
73	100		
57	22.5	18.8	28.2




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#20
Methylene Chloride
Concen: 43.973 ug/l
RT: 2.84 min Scan# 321
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
84	100		
49	124.2	95.8	143.6
51	38.1	29.8	44.8
86	64.1	50.8	76.2





#21

trans-1,2-Dichloroethene

Concen: 45.341 ug/l

RT: 3.15 min Scan# 372

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	96	Resp:	312681
Ion Ratio	Lower	Upper	
96	100		
61	138.3	108.3	162.5
98	64.9	50.8	76.2

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#22

Diisopropyl ether

Concen: 47.289 ug/l




RT: 3.84 min Scan# 484


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	45	Resp:	1053210
Ion Ratio	Lower	Upper	
45	100		
43	75.1	57.4	86.0
87	26.1	21.9	32.9
59	11.4	9.0	13.6





#23

Vinyl Acetate

Concen: 247.256 ug/l

RT: 3.80 min Scan# 478

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 4493937

Ion Ratio Lower Upper

43 100





86 10.8 8.6 12.8

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#24

1,1-Dichloroethane

Concen: 46.291 ug/l

RT: 3.69 min Scan# 459

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02




Tgt Ion: 63 Resp: 573601


Ion Ratio Lower Upper

63 100

98 7.7 3.6 10.8

100 4.7 2.3 6.8





#25

2-Butanone

Concen: 243.775 ug/l

RT: 4.65 min Scan# 618

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 1439048

Ion Ratio Lower Upper

43 100





72 26.0 21.0 31.4

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#26

2,2-Dichloropropane

Concen: 49.343 ug/l

RT: 4.56 min Scan# 603

Delta R.T. -0.01 min

Lab File: VX014023.D




Acq: 16 Dec 2019 10:02


Tgt Ion: 77 Resp: 474716

Ion Ratio Lower Upper

77 100

97 23.7 11.9 35.9





#27





cis-1,2-Dichloroethene
Concen: 46.255 ug/l
RT: 4.58 min Scan# 606
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
96	100			
61	143.9	0.0	288.4	
98	64.4	0.0	129.6	

Manual Integrations
APPROVED




apatel
12/17/2019 4:50:04 PM




#28

Bromochloromethane
Concen: 52.042 ug/l
RT: 5.00 min Scan# 674
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Resp	Lower	Upper
49	100			
129	2.4	0.0	5.0	
130	78.0	64.6	97.0	









#29
Tetrahydrofuran
Concen: 229.867 ug/l
RT: 5.11 min Scan# 692
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
42	100		
72	43.9	35.8	53.8
71	40.5	33.6	50.4




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#30
Chloroform
Concen: 46.877 ug/l
RT: 5.20 min Scan# 707
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
83	100		
85	66.6	50.8	76.2









#31
Cyclohexane
Concen: 48.243 ug/l
RT: 5.57 min Scan# 768
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
56	100		
69	32.0	23.2	34.8
84	88.3	69.2	103.8




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#32
1,1,1-Trichloroethane
Concen: 46.872 ug/l
RT: 5.48 min Scan# 753
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
97	100		
99	64.5	52.0	78.0
61	45.1	36.7	55.1









#33
 1,2-Dichloroethane-d4
 Concen: 47.252 ug/l
 RT: 6.04 min Scan# 846
 Delta R.T. -0.01 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
65	100			
67	53.2	0.0	106.4	




Manual Integrations
 APPROVED


apatel
 12/17/2019 4:50:04 PM



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.84 min Scan# 977
 Delta R.T. -0.01 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Resp	Lower	Upper
114	100			
63	19.6	0.0	40.8	
88	15.9	0.0	30.4	





#35

Dibromofluoromethane

Concen: 47.899 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	113	Resp:	284634
Ion Ratio	Lower	Upper	
113	100		
111	102.8	82.0	123.0
192	22.9	19.3	28.9

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#36

1,1-Dichloropropene

Concen: 47.956 ug/l




RT: 5.79 min Scan# 804


Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	75	Resp:	430143
Ion Ratio	Lower	Upper	
75	100		
110	36.0	18.3	54.9
77	30.6	24.8	37.2









#37
Ethyl Acetate
Concen: 46.954 ug/l
RT: 4.81 min Scan# 644
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	43	Resp	465070
Ion Ratio	Lower	Upper	
43	100		
61	13.5	10.8	16.2
70	10.6	8.6	12.8




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#38
Carbon Tetrachloride
Concen: 49.479 ug/l
RT: 5.77 min Scan# 801
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	117	Resp	404296
Ion Ratio	Lower	Upper	
117	100		
119	94.6	76.2	114.4
121	30.4	23.6	35.4









#39
Methvlcvclohexane
Concen: 49.367 ug/l
RT: 7.45 min Scan# 1077
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
83	100	546743		
55	78.3	61.0	91.6	
98	48.8	38.6	57.8	




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#40
Benzene
Concen: 46.948 ug/l
RT: 6.13 min Scan# 860
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Resp	Lower	Upper
78	100	1295678		
77	23.4	18.8	28.2	





#41

Methacrylonitrile

Concen: 47.349 ug/l

RT: 5.02 min Scan# 678

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion	Ratio	Lower	Upper
41	100		
39	54.8	44.5	66.7
67	69.1	57.4	86.0
52	27.5	23.0	34.4

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#42

1,2-Dichloroethane

Concen: 47.582 ug/l




RT: 6.18 min Scan# 868


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	21.0





#43

Isopropyl Acetate

Concen: 47.240 ug/l

RT: 6.43 min Scan# 909

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion	43	Resp	774355
Ion Ratio	Lower	Upper	
43	100		
61	20.5	16.4	24.6
87	13.1	10.7	16.1

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#44

Trichloroethene

Concen: 45.631 ug/l




RT: 7.20 min Scan# 1036


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion	130	Resp	347833
Ion Ratio	Lower	Upper	
130	100		
95	93.1	0.0	185.6





#45

1,2-Dichloropropane

Concen: 47.792 ug/l

RT: 7.50 min Scan# 1085

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 63 Resp: 338061

Ion Ratio Lower Upper

63 100





65 31.4 25.8 38.8

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#46

Dibromomethane

Concen: 46.498 ug/l

RT: 7.65 min Scan# 1109

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02




Tgt Ion: 93 Resp: 217459


Ion Ratio Lower Upper

93 100

95 82.7 67.3 100.9

174 112.6 91.6 137.4





#47

Bromodichloromethane

Concen: 49.805 ug/l

RT: 7.89 min Scan# 1148

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	83	Resp:	438682
Ion Ratio	Lower	Upper	
83	100		
85	64.7	51.8	77.8
127	8.2	7.0	10.4

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#48

Methyl methacrylate

Concen: 47.974 ug/l




RT: 7.76 min Scan# 1127


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	41	Resp:	385076
Ion Ratio	Lower	Upper	
41	100		
69	80.2	65.8	98.6
39	55.6	44.6	67.0





#49

1,4-Dioxane

Concen: 871.049 ug/l

RT: 7.73 min Scan# 1122

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	88	Resp:	142282
Ion Ratio	Lower	Upper	
88	100		
43	33.8	26.5	39.7
58	70.2	56.8	85.2

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#50

Toluene-d8

Concen: 48.003 ug/l




RT: 8.71 min Scan# 1283


Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	98	Resp:	1110544
Ion Ratio	Lower	Upper	
98	100		
100	66.1	52.9	79.3





#51

4-Methyl-2-Pentanone

Concen: 229.864 ug/l

RT: 8.64 min Scan# 1271

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 2366328

Ion Ratio Lower Upper

43 100


58 39.9 32.2 48.2

Manual Integrations

APPROVED




apatel

12/17/2019 4:50:04 PM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 46.505 ug/l

RT: 8.78 min Scan# 1294

Delta R.T. -0.01 min

Lab File: VX014023.D


Acq: 16 Dec 2019 10:02

Tgt Ion: 92 Resp: 811227

Ion Ratio Lower Upper



92 100


91 171.5 136.2 204.4



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53
t-1,3-Dichloropropene
Concen: 50.984 ug/l
RT: 9.03 min Scan# 1336
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 75 Resp: 492403
Ion Ratio Lower Upper
75 100
77 32.5 25.1 37.7




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#54
cis-1,3-Dichloropropene
Concen: 51.088 ug/l
RT: 8.43 min Scan# 1237
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion: 75 Resp: 549690
Ion Ratio Lower Upper
75 100
77 31.4 25.3 37.9
39 49.6 39.9 59.9









#55
 1,1,2-Trichloroethane
 Concen: 46.216 ug/l
 RT: 9.21 min Scan# 1365
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ion	Ratio	Resp	Lower	Upper
97	100		323112		
83	85.4	68.2	102.4		
85	55.6	44.6	66.8		
99	61.0	51.4	77.0		




Manual Integrations
 APPROVED


apatel
 12/17/2019 4:50:04 PM



#56
 Ethyl methacrylate
 Concen: 48.331 ug/l
 RT: 9.17 min Scan# 1359
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
69	100		530780		
41	68.7	54.8	82.2		
39	36.0	28.3	42.5		





#57

1,3-Dichloropropane

Concen: 46.663 ug/l

RT: 9.36 min Scan# 1390

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 76 Resp: 548890

Ion Ratio Lower Upper

76 100





78 32.0 25.8 38.6

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#58

2-Chloroethyl Vinyl ether

Concen: 264.882 ug/l

RT: 8.30 min Scan# 1216

Delta R.T. -0.01 min

Lab File: VX014023.D




Acq: 16 Dec 2019 10:02


Tgt Ion: 63 Resp: 1100864

Ion Ratio Lower Upper

63 100

106 28.5 23.0 34.6









#59
2-Hexanone
Concen: 243.985 ug/l
RT: 9.48 min Scan# 1410
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 43 Resp: 2021919
Ion Ratio Lower Upper
43 100
58 54.8 28.0 84.0




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#60
Dibromochloromethane
Concen: 50.848 ug/l
RT: 9.57 min Scan# 1425
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion: 129 Resp: 353367
Ion Ratio Lower Upper
129 100
127 76.6 38.4 115.2





#61

1,2-Dibromoethane

Concen: 48.089 ug/l

RT: 9.67 min Scan# 1440

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 107 Resp: 343988

Ion Ratio Lower Upper


107 100

109 93.5 75.7 113.5

Manual Integrations
APPROVED




apatel

12/17/2019 4:50:04 PM



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 46.691 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02


Tgt Ion: 95 Resp: 394862

Ion Ratio Lower Upper

95 100

174 88.9 0.0 175.8



176 85.3 0.0 173.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1512
Delta R.T. -0.01 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
117	100		
82	54.7	42.2	63.4
119	31.1	25.1	37.7




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#64
Tetrachloroethene
Concen: 43.312 ug/l
RT: 9.33 min Scan# 1385
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
164	100		
166	132.7	104.0	156.0
129	89.7	72.2	108.4
131	88.1	69.6	104.4









#65
Chlorobenzene
Concen: 46.388 ug/l
RT: 10.14 min Scan# 1517
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion:112 Resp: 865726
Ion Ratio Lower Upper
112 100
114 32.0 24.9 37.3




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 48.110 ug/l
RT: 10.21 min Scan# 1530
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion:131 Resp: 320595
Ion Ratio Lower Upper
131 100
133 95.4 48.0 144.0
119 66.1 33.4 100.2









#67
Ethyl Benzene
Concen: 48.500 ug/l
RT: 10.25 min Scan# 1535
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 91 Resp: 1556159
Ion Ratio Lower Upper
91 100
106 31.5 25.0 37.6




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#68
m/p-Xylenes
Concen: 96.095 ug/l
RT: 10.35 min Scan# 1553
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion: 106 Resp: 1185875
Ion Ratio Lower Upper
106 100
91 198.7 158.6 238.0









#69
o-Xylene
Concen: 47.071 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
106	100		
91	212.4	104.2	312.6




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#70
Styrene
Concen: 48.371 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
104	100		
78	49.1	38.5	57.7
103	54.7	42.9	64.3





#71

Bromoform

Concen: 50.341 ug/l

RT: 10.85 min Scan# 1634

Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	173	Resp:	269651
Ion Ratio	Lower	Upper	
173	100		
175	49.2	24.4	73.4
254	0.2	0.2	0.2

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.07 min Scan# 1835


Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	152	Resp:	439040
Ion Ratio	Lower	Upper	
152	100		
115	77.9	38.3	114.9
150	174.2	0.0	345.4





#73

Isopropylbenzene

Concen: 48.908 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 105 Resp: 1511830

Ion Ratio Lower Upper

105 100





120 27.3 13.5 40.4

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#74

N-ethyl acetate

Concen: 48.883 ug/l

RT: 10.89 min Scan# 1641

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion: 43 Resp: 708291




Ion Ratio Lower Upper


43 100

70 42.4 34.4 51.6

55 23.4 19.1 28.7

61 24.5 19.7 29.5





#75

1,1,2,2-Tetrachloroethane

Concen: 47.031 ug/l

RT: 11.26 min Scan# 1702

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	83	Resp:	504024
Ion Ratio		Lower	Upper
83	100		
131	10.0	5.1	15.2
85	64.0	31.9	95.7

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#76

1,2,3-Trichloropropane

Concen: 42.287 ug/l m




RT: 11.29 min Scan# 1706


Delta R.T. -0.01 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	75	Resp:	403054
Ion Ratio		Lower	Upper
75	100		
77	44.7	19.3	57.8





#77

Bromobenzene

Concen: 46.144 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	156	Resp:	392452
Ion Ratio	Lower	Upper	
156	100		
77	153.4	76.5	229.5
158	96.9	49.3	147.9

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#78

n-propylbenzene

Concen: 50.339 ug/l




RT: 11.35 min Scan# 1717


Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	91	Resp:	1745388
Ion Ratio	Lower	Upper	
91	100		
120	23.3	11.7	35.0









#79
 2-Chlorotoluene
 Concen: 48.371 ug/l
 RT: 11.42 min Scan# 1727
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion: 91 Resp: 1019503
 Ion Ratio Lower Upper
 91 100
 126 34.5 17.2 51.6




Manual Integrations
 APPROVED


apatel
 12/17/2019 4:50:04 PM



#80
 1,3,5-Trimethylbenzene
 Concen: 48.835 ug/l
 RT: 11.50 min Scan# 1741
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion: 105 Resp: 1270462
 Ion Ratio Lower Upper
 105 100
 120 50.5 25.3 75.8









#81
trans-1,4-Dichloro-2-butene
Concen: 51.939 ug/l
RT: 11.07 min Scan# 1670
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
75	100		
53	95.7	76.7	115.1
89	46.1	34.6	51.8




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#82
4-Chlorotoluene
Concen: 48.359 ug/l
RT: 11.51 min Scan# 1742
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
91	100		
126	30.4	15.6	46.8









#83
tert-Butylbenzene
Concen: 49.723 ug/l
RT: 11.76 min Scan# 1784
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
119	100		
91	55.2	28.5	85.6
134	23.6	12.2	36.6




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#84
1,2,4-Trimethylbenzene
Concen: 49.181 ug/l
RT: 11.80 min Scan# 1790
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
105	100		
120	46.0	23.1	69.2





#85

sec-Butylbenzene

Concen: 49.742 ug/l

RT: 11.94 min Scan# 1813

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion:105 Resp: 1486682

Ion Ratio Lower Upper

105 100





134 20.5 10.4 31.1

Manual Integrations

APPROVED

apatel

12/17/2019 4:50:04 PM



#86

p-Isopropyltoluene

Concen: 50.421 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02




Tgt Ion:119 Resp: 1378464


Ion Ratio Lower Upper

119 100

134 26.5 13.4 40.1

91 22.5 11.4 34.1









#87
 1,3-Dichlorobenzene
 Concen: 46.468 ug/l
 RT: 12.02 min Scan# 1826
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	38.0	19.1	57.1
148	63.7	32.3	96.9




Manual Integrations
 APPROVED


apatel
 12/17/2019 4:50:04 PM



#88
 1,4-Dichlorobenzene
 Concen: 45.688 ug/l
 RT: 12.09 min Scan# 1838
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
146	100		
111	37.8	18.7	56.1
148	63.6	31.9	95.9









#89
n-Butylbenzene
Concen: 51.572 ug/l
RT: 12.39 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
91	100		
92	54.8	27.2	81.6
134	26.9	13.4	40.1




Manual Integrations
APPROVED


apatel
12/17/2019 4:50:04 PM



#90
Hexachloroethane
Concen: 51.121 ug/l
RT: 12.59 min Scan# 1920
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
117	100		
201	85.0	43.1	129.3









#91
 1,2-Dichlorobenzene
 Concen: 46.266 ug/l
 RT: 12.39 min Scan# 1886
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	39.3	19.7	59.1
148	63.1	32.1	96.5




Manual Integrations
 APPROVED


apatel
 12/17/2019 4:50:04 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 44.248 ug/l
 RT: 12.99 min Scan# 1986
 Delta R.T. 0.00 min
 Lab File: VX014023.D
 Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
75	100		
155	97.3	46.9	140.6
157	121.9	60.8	182.4





#93

1,2,4-Trichlorobenzene

Concen: 49.178 ug/l

RT: 13.64 min Scan# 2092

Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	180	Resp:	479619
Ion Ratio	Lower	Upper	
180	100		
182	95.0	46.6	139.8
145	27.9	14.2	42.6

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#94

Hexachlorobutadiene

Concen: 47.397 ug/l




RT: 13.78 min Scan# 2114


Delta R.T. 0.00 min

Lab File: VX014023.D

Acq: 16 Dec 2019 10:02

Tgt Ion:	225	Resp:	222154
Ion Ratio	Lower	Upper	
225	100		
223	64.7	30.9	92.5
227	63.4	30.9	92.7









#95
Naphthalene
Concen: 49.867 ug/l
RT: 13.83 min Scan# 2123
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.2	15.4
129	10.7	8.7	13.1

Manual Integrations
APPROVED

apatel
12/17/2019 4:50:04 PM



#96
1,2,3-Trichlorobenzene
Concen: 48.439 ug/l
RT: 14.01 min Scan# 2153
Delta R.T. 0.00 min
Lab File: VX014023.D
Acq: 16 Dec 2019 10:02

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.8	46.8	140.3
145	30.4	14.8	44.4

