

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025815.D
 Acq On : 15 Dec 2021 15:28
 Operator : JC/MD
 Sample : M4988-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20211206DL

Quant Time: Dec 16 07:29:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	268915	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103626	46.510	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.020%
35) Dibromofluoromethane	5.391	113	85936	45.822	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.640%
50) Toluene-d8	8.653	98	311310	45.558	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.120%#
62) 4-Bromofluorobenzene	11.085	95	117958	45.444	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.880%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	4399	2.668	ug/l	96
12) 1,1-Dichloroethene	2.318	96	1059	0.646	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	1310	0.604	ug/l	97
44) Trichloroethene	7.129	130	98579	46.672	ug/l	98
64) Tetrachloroethene	9.275	164	2062	1.070	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025815.D
Acq On : 15 Dec 2021 15:28
Operator : JC/MD
Sample : M4988-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20211206DL

Quant Time: Dec 16 07:29:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

