

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025818.D
 Acq On : 15 Dec 2021 16:39
 Operator : JC/MD
 Sample : M4988-14 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20211207

Quant Time: Dec 16 07:30:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95422	47.259	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.520%		
35) Dibromofluoromethane	5.391	113	78464	46.169	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.340%		
50) Toluene-d8	8.653	98	289803	46.801	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.600%		
62) 4-Bromofluorobenzene	11.085	95	109730	46.650	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.300%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	535	0.358	ug/l #	72
44) Trichloroethene	7.129	130	66272	34.625	ug/l	95
64) Tetrachloroethene	9.281	164	369	0.209	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025818.D
Acq On : 15 Dec 2021 16:39
Operator : JC/MD
Sample : M4988-14 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20211207

Quant Time: Dec 16 07:30:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

