

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025820.D
 Acq On : 15 Dec 2021 17:26
 Operator : JC/MD
 Sample : PB141382TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141382TB

Quant Time: Dec 16 07:30:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96698	47.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.700%
35) Dibromofluoromethane	5.391	113	79637	46.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.700%
50) Toluene-d8	8.653	98	296186	47.312	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.620%
62) 4-Bromofluorobenzene	11.085	95	103969	43.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%
Target Compounds						
					Qvalue	
25) 2-Butanone	4.568	43	4650	3.421	ug/l #	86
43) Isopropyl Acetate	6.348	43	12434	2.981	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025820.D
Acq On : 15 Dec 2021 17:26
Operator : JC/MD
Sample : PB141382TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141382TB

Quant Time: Dec 16 07:30:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

