

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025823.D
 Acq On : 15 Dec 2021 18:37
 Operator : JC/MD
 Sample : PB141382ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141382ZHE#03

Quant Time: Dec 16 07:31:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238674	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94210	48.591	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.180%		
35) Dibromofluoromethane	5.391	113	77023	46.273	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.540%		
50) Toluene-d8	8.653	98	289041	47.659	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.320%		
62) 4-Bromofluorobenzene	11.085	95	96711	41.979	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.960%		
Target Compounds						
25) 2-Butanone	4.574	43	4454	3.416	ug/l	96
43) Isopropyl Acetate	6.342	43	14197	3.514	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025823.D
Acq On : 15 Dec 2021 18:37
Operator : JC/MD
Sample : PB141382ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141382ZHE#03

Quant Time: Dec 16 07:31:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

