

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025829.D
 Acq On : 15 Dec 2021 20:57
 Operator : JC/MD
 Sample : M5074-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-2

Quant Time: Dec 16 07:32:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242214	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94579	47.192	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.380%
35) Dibromofluoromethane	5.391	113	78824	46.663	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.653	98	295819	48.064	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.120%
62) 4-Bromofluorobenzene	11.085	95	114602	49.018	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.040%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	8636	2.752	ug/l	96
25) 2-Butanone	4.580	43	4728	3.508	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025829.D
Acq On : 15 Dec 2021 20:57
Operator : JC/MD
Sample : M5074-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-2

Quant Time: Dec 16 07:32:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

