

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025831.D
 Acq On : 15 Dec 2021 21:44
 Operator : JC/MD
 Sample : M5067-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BRICK

Quant Time: Dec 16 07:32:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	250712	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100745	48.146	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.300%
35) Dibromofluoromethane	5.391	113	81703	46.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.460%
50) Toluene-d8	8.653	98	305080	47.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.780%
62) 4-Bromofluorobenzene	11.085	95	113473	46.890	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.780%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	6165	1.078	ug/l	98
25) 2-Butanone	4.568	43	4755	3.379	ug/l	94
43) Isopropyl Acetate	6.348	43	11227	2.645	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025831.D
Acq On : 15 Dec 2021 21:44
Operator : JC/MD
Sample : M5067-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BRICK

Quant Time: Dec 16 07:32:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

