

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025832.D
 Acq On : 15 Dec 2021 22:08
 Operator : JC/MD
 Sample : M5067-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BASE

Quant Time: Dec 16 07:32:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236130	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94160	48.716	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.440%		
35) Dibromofluoromethane	5.391	113	76775	46.621	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.240%		
50) Toluene-d8	8.653	98	283775	47.295	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.580%		
62) 4-Bromofluorobenzene	11.085	95	103900	45.586	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.180%		
Target Compounds						
20) Methylene Chloride	2.794	84	6621	1.666	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025832.D
Acq On : 15 Dec 2021 22:08
Operator : JC/MD
Sample : M5067-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BASE

Quant Time: Dec 16 07:32:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

