

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025839.D
 Acq On : 16 Dec 2021 02:02
 Operator : JC/MD
 Sample : M5084-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C1

Quant Time: Dec 16 07:35:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142311	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93855	45.837	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.680%
35) Dibromofluoromethane	5.391	113	78816	45.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.653	98	287196	45.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.440%#
62) 4-Bromofluorobenzene	11.085	95	104547	43.819	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.640%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.514	76	1115	0.258	ug/l	97
39) Methylcyclohexane	7.385	83	9504	3.633	ug/l	97
73) Isopropylbenzene	10.963	105	151756	20.329	ug/l	100
78) n-propylbenzene	11.305	91	245703	29.099	ug/l	99
83) tert-Butylbenzene	11.719	119	8790	1.448	ug/l	85
85) sec-Butylbenzene	11.890	105	77838	10.671	ug/l	90
89) n-Butylbenzene	12.335	91	41201	8.063	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025839.D
Acq On : 16 Dec 2021 02:02
Operator : JC/MD
Sample : M5084-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C1

Quant Time: Dec 16 07:35:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

