

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025840.D
 Acq On : 16 Dec 2021 02:25
 Operator : JC/MD
 Sample : M5084-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C2

Quant Time: Dec 16 07:35:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97507	44.078	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.160%
35) Dibromofluoromethane	5.397	113	82683	46.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.060%
50) Toluene-d8	8.653	98	298314	45.583	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.160%#
62) 4-Bromofluorobenzene	11.085	95	105767	42.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.100%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	1383	0.297	ug/l	96
39) Methylcyclohexane	7.373	83	608	0.223	ug/l #	79
73) Isopropylbenzene	10.963	105	27163	3.639	ug/l	100
78) n-propylbenzene	11.305	91	15482	1.834	ug/l	98
83) tert-Butylbenzene	11.713	119	4040	0.666	ug/l	92
85) sec-Butylbenzene	11.890	105	26974	3.699	ug/l	89
89) n-Butylbenzene	12.335	91	8302	1.625	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025840.D
Acq On : 16 Dec 2021 02:25
Operator : JC/MD
Sample : M5084-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C2

Quant Time: Dec 16 07:35:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

