

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025843.D
 Acq On : 16 Dec 2021 03:36
 Operator : JC/MD
 Sample : M5084-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C6

Quant Time: Dec 16 07:35:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96750	43.608	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.220%
35) Dibromofluoromethane	5.391	113	73732	40.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	81.940%
50) Toluene-d8	8.653	98	300254	45.796	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.600%#
62) 4-Bromofluorobenzene	11.085	95	100565	40.379	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.760%#
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	54729	39.347	ug/l	98
25) 2-Butanone	4.562	43	16754	11.228	ug/l	91
30) Chloroform	5.099	83	9955	2.751	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025843.D
Acq On : 16 Dec 2021 03:36
Operator : JC/MD
Sample : M5084-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C6

Quant Time: Dec 16 07:35:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

