

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025844.D
 Acq On : 16 Dec 2021 03:59
 Operator : JC/MD
 Sample : M5084-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C7

Quant Time: Dec 16 07:36:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254554	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97960	46.250	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.500%
35) Dibromofluoromethane	5.391	113	80489	45.339	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.680%
50) Toluene-d8	8.653	98	293263	45.339	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.680%#
62) 4-Bromofluorobenzene	11.085	95	103956	42.309	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.620%
Target Compounds						
					Qvalue	
73) Isopropylbenzene	10.963	105	9872	1.367	ug/l	100
83) tert-Butylbenzene	11.719	119	4066	0.692	ug/l	88
85) sec-Butylbenzene	11.890	105	10907	1.545	ug/l	81
89) n-Butylbenzene	12.335	91	1502	0.304	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
Data File : VX025844.D
Acq On : 16 Dec 2021 03:59
Operator : JC/MD
Sample : M5084-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C7

Quant Time: Dec 16 07:36:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

