


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121621\
 Data File : VX025809.D
 Acq On : 15 Dec 2021 12:34
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Title : SW846 8260
 Last Update : Tue Dec 14 07:02:11 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	5895	PASS
75	95	30	60	49.5	15249	PASS
95	95	100	100	100.0	30779	PASS
96	95	5	9	6.4	1968	PASS
173	174	0.00	2	1.2	284	PASS
174	95	50	100	77.9	23979	PASS
175	174	5	9	6.9	1654	PASS
176	174	95	101	97.7	23437	PASS
177	176	5	9	6.7	1568	PASS