

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033451.D
 Acq On : 16 Dec 2022 10:53
 Operator : JC/MD
 Sample : VX1216WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1216WBL01

Quant Time: Dec 17 07:44:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Dec 12 12:37:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	12863	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	90341	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	96718	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	35806	31.695	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.667%
60) 4-Bromofluorobenzene	11.079	95	49264	27.502	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.667%
63) Toluene-d8	8.647	98	95028	29.536	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.467%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
Data File : VX033451.D
Acq On : 16 Dec 2022 10:53
Operator : JC/MD
Sample : VX1216WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1216WBL01

Quant Time: Dec 17 07:44:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Dec 12 12:37:35 2022
Response via : Initial Calibration

