

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033454.D
 Acq On : 16 Dec 2022 12:16
 Operator : JC/MD
 Sample : N6074-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 221214021-01-TRIP-BLANK

Quant Time: Dec 17 07:45:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Dec 12 12:37:35 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/19/2022
 Supervised By : Mahesh Dadoda 12/19/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	8661	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	81258	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	85973	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	32011	42.083	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	140.267%#
60) 4-Bromofluorobenzene	11.079	95	43553	27.352	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.167%
63) Toluene-d8	8.647	98	84149	29.424	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.067%
Target Compounds						
15) Acetone	2.392	58	1089	8.281	ug/l	Qvalue 75
25) Chloroform	5.111	83	1438m	0.831	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
Data File : VX033454.D
Acq On : 16 Dec 2022 12:16
Operator : JC/MD
Sample : N6074-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
221214021-01-TRIP-BLANK

Quant Time: Dec 17 07:45:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X121222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Dec 12 12:37:35 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/19/2022
Supervised By :Mahesh Dadoda 12/19/2022

