

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033461.D
 Acq On : 16 Dec 2022 15:20
 Operator : JC/MD
 Sample : VX1216WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1216WBL02

Quant Time: Dec 17 08:09:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	198269	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	35919	43.003	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.000%
35) Dibromofluoromethane	5.385	113	41077	46.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.120%
50) Toluene-d8	8.647	98	84875	45.496	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.000%#
62) 4-Bromofluorobenzene	11.079	95	65386	45.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
Data File : VX033461.D
Acq On : 16 Dec 2022 15:20
Operator : JC/MD
Sample : VX1216WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1216WBL02

Quant Time: Dec 17 08:09:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

