

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033465.D
 Acq On : 16 Dec 2022 17:00
 Operator : JC/MD
 Sample : N6089-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-24

Quant Time: Dec 17 08:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38168	46.271	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.540%
35) Dibromofluoromethane	5.391	113	42994	46.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.653	98	91573	47.548	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.100%
62) 4-Bromofluorobenzene	11.079	95	71564	48.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.200%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18954	31.052	ug/l	99
20) Methylene Chloride	2.788	84	11348	8.336	ug/l	90
37) Ethyl Acetate	4.721	43	15873	9.355	ug/l	99
43) Isopropyl Acetate	6.342	43	73124	26.785	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
Data File : VX033465.D
Acq On : 16 Dec 2022 17:00
Operator : JC/MD
Sample : N6089-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-24

Quant Time: Dec 17 08:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

