

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
 Data File : VX033468.D
 Acq On : 16 Dec 2022 18:09
 Operator : JC/MD
 Sample : N6071-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FVSUB-TP8-COMP-08

Quant Time: Dec 17 08:12:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	125796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39974	49.533	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.060%
35) Dibromofluoromethane	5.385	113	44042	48.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	8.653	98	93349	48.780	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.560%
62) 4-Bromofluorobenzene	11.079	95	72676	49.668	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.340%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	24381	40.827	ug/l	99
20) Methylene Chloride	2.794	84	8230	6.180	ug/l #	92
37) Ethyl Acetate	4.715	43	14673	8.703	ug/l	98
43) Isopropyl Acetate	6.336	43	65712	24.224	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121622\
Data File : VX033468.D
Acq On : 16 Dec 2022 18:09
Operator : JC/MD
Sample : N6071-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FVSUB-TP8-COMP-08

Quant Time: Dec 17 08:12:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

